

§ 95.1221 RF exposure.

A MedRadio medical implant device or medical body-worn transmitter is subject to the radiofrequency radiation exposure requirements specified in §§ 1.1307(b) and 2.1093 of this chapter, as appropriate. Applications for equipment authorization of devices operating under this section must demonstrate compliance with these requirements using either finite difference time domain (FDTD) computational modeling or laboratory measurement techniques. Where a showing is based on computational modeling, the Commission retains the discretion to request that supporting documentation and/or specific absorption rate (SAR) measurement data be submitted.

[78 FR 33653, June 4, 2013]

§ 95.1223 Registration and frequency coordination.

(a) *Registration.* Prior to operating MBAN devices that are capable of operation in the 2360–2390 MHz band, a health care facility, as defined by § 95.1203, must register with a frequency coordinator designated under § 95.1225. Operation of MBAN devices in the 2360–2390 MHz band is prohibited prior to the MBAN coordinator notifying the health care facility that registration and coordination (to the extent coordination is required under paragraph (c) of this section) is complete. The registration must include the following information:

(1) Specific frequencies or frequency range(s) within the 2360–2390 MHz band to be used, and the capabilities of the MBAN equipment to use the 2390–2400 MHz band;

(2) Effective isotropic radiated power;

(3) Number of MedRadio programmer/control transmitters in use at the health care facility as of the date of registration including manufacturer name(s) and model numbers and FCC identification number;

(4) Legal name of the health care facility;

(5) Location of MedRadio programmer/control transmitters (*e.g.*, geographic coordinates, street address, building);

(6) Point of contact for the health care facility (*e.g.*, name, title, office,

phone number, fax number, email address); and

(7) In the event an MBAN has to cease operating in all or a portion of the 2360–2390 MHz band due to interference under § 95.1211 or changes in coordination under paragraph (c) of this section, a point of contact (including contractors) for the health care facility that is responsible for ensuring that this change is effected whenever it is required (*e.g.*, name, title, office, phone number, fax number, email address). The health care facility also must state whether, in such cases, its MBAN operation is capable of defaulting to the 2390–2400 MHz band and that it is responsible for ceasing MBAN operations in the 2360–2390 MHz band or defaulting traffic to other hospital systems.

(b) *Notification.* A health care facility shall notify the frequency coordinator whenever an MBAN programmer/control transmitter in the 2360–2390 MHz band is permanently taken out of service, unless it is replaced with transmitter(s) using the same technical characteristics and locations as those reported on the health care facility's registration which will cover the replacement transmitter(s). A health care facility shall keep the information contained in each registration current and shall notify the frequency coordinator of any material change to the MBAN's location or operating parameters. In the event that the health care facility proposes to change the MBAN's location or operating parameters, the MBAN coordinator must first evaluate the proposed changes and comply with paragraph (c) of this section, as appropriate, before the health care facility may operate the MBAN in the 2360–2390 MHz band under changed operating parameters

(c) *Coordination procedures.* The frequency coordinator will determine if an MBAN is within the line of sight of an AMT receive facility in the 2360–2390 MHz band and notify the health care facility when it may begin MBAN operations under the applicable procedures in (c)(1) or (2) of this section.

(1) If the MBAN is beyond the line of sight of an AMT receive facility, it

§ 95.1225

47 CFR Ch. I (10–1–16 Edition)

may operate without prior coordination with the AMT coordinator, provided that the MBAN coordinator provides the AMT coordinator with the MBAN registration information and the AMT coordinator concurs that the MBAN is beyond the line of sight prior to the MBAN beginning operations in the band.

(2) If the MBAN is within line of sight of an AMT receive facility, the MBAN frequency coordinator shall achieve a mutually satisfactory coordination agreement with the AMT frequency coordinator prior to the MBAN beginning operations in the band. Such coordination agreement shall provide protection to AMT receive stations consistent with International Telecommunication Union (ITU) Recommendation ITU-R M.1459, “Protection criteria for telemetry systems in the aeronautical mobile service and mitigation techniques to facilitate sharing with geostationary broadcasting-satellite and mobile-satellite services in the frequency bands 1 452–1 525 and 2 310–2 360 MHz,” May 2000, as adjusted using generally accepted engineering practices and standards that are mutually agreeable to both coordinators to take into account the local conditions and operating characteristics of the applicable AMT and MBAN facilities, and shall specify when the device shall limit its transmissions to segments of the 2360–2390 MHz band or shall cease operation in the band. This ITU document is incorporated by reference in accordance with 5 U.S.C. 552(a) and 1 CFR part 51 and approved by the Director of Federal Register. Copies of the recommendation may be obtained from ITU, Place des Nations, 1211 Geneva 20, Switzerland, or online at <http://www.itu.int/en/publications/Pages/default.aspx>. You may inspect a copy at the Federal Communications Commission, 445 12th Street, SW., Washington, DC 20554, or at the National Archives and Records Administration (NARA). For information on the availability of this material at NARA, call 202-741-6030, or go to: http://www.archives.gov/federal_register/code_of_federal_regulations/ibr_locations.html. “Generally accepted engineering practices and standards” include, but are not limited to, engi-

neering analyses and measurement data as well as limiting MBAN operations in the band by time or frequency.

(3) If an AMT operator plans to operate a receive site not previously analyzed by the MBAN coordinator to determine line of sight to an MBAN facility, the AMT operator shall consider using locations that are beyond the line of sight of a registered health care facility. If the AMT operator determines that non-line of sight locations are not practical for its purposes, the AMT coordinator shall notify the MBAN coordinator upon no less than 7 days’ notice that the registered health care facility must cease MBAN operations in the 2360–2390 MHz band unless the parties can achieve a mutually satisfactory coordination agreement under paragraph (c)(2) of this section.

[77 FR 55734, Sept. 11, 2012, as amended at 79 FR 60100, Oct. 6, 2014]

§ 95.1225 Frequency coordinator.

(a) The Commission will designate a frequency coordinator(s) to manage the operation of medical body area networks by eligible health care facilities.

(b) The frequency coordinator shall perform the following functions:

(1) Register health care facilities that operate MBAN transmitters, maintain a database of these MBAN transmitter locations and operational parameters, and provide the Commission with information contained in the database upon request;

(2) Determine if an MBAN is within line of sight of an AMT receive facility in the 2360–2390 MHz band and coordinate MBAN operations with the designated AMT coordinator as specified in § 87.305 of this chapter;

(3) Notify a registered health care facility when an MBAN has to change frequency within the 2360–2390 MHz band or to cease operating in the band consistent with a coordination agreement between the MBAN and the AMT coordinators;

(4) Develop procedures to ensure that registered health care facilities operate an MBAN consistent with the coordination requirements under § 95.1223; and

(5) Identify the MBAN that is the source of interference in response to a complaint from the AMT coordinator